

The temperature sensor is connected to the bottom of the 7500 unit, which is the white box just to the left of the tower. For flux calculations, one should not use the temperature from this instrument, but rather the virtual temperature (T_s in the data file) from the sonic.

